

Air Monitoring Summary Tables

The table below summarizes monitoring data collected using the H₂S analyzers deployed at the onsite stations.
All times in Eastern Standard Time (EST).

From: 11/21/24 12:00 am

To: 11/21/24 11:59 pm

Onsite Fenceline Monitors

Instrument	Analyte	30-min AEGL Reached?	Concentration Range Detected ^a	24-hr Average ^a	7-day Average	30-min AEGL
Station 1						
TAPI Analyzer	H ₂ S	No	0 – 2 ppb	0.69 ppb ^b	1.42 ppb	600 ppb
Station 2						
TAPI Analyzer	H ₂ S	No	0 – 3 ppb	0.76 ppb ^c	0.45 ppb	600 ppb
Station 3						
TAPI Analyzer	H ₂ S	No	0 – 3 ppb	0.60 ppb	0.61 ppb	600 ppb

^a Based on 30-minute averages.


^b The 24-hour H₂S average at Station 1 from the start of the sampling period to the end of the sampling period is represented by the backup unit data.

^c The 24-hour H₂S average at Station 2 from the start of the sampling period to the end of the sampling period is represented by the backup unit data

Notes:

ATSDR MRL	Agency for Toxic Substances and Disease Registry Minimal Risk Level (MRL)
AEGL	EPA Acute Exposure Guidelines Levels
H ₂ S	Hydrogen Sulfide
TAPI	Teledyne API H ₂ S Analyzer
hr	Hour
min	Minute
ppb	Parts per billion
MRL Limit	Limit defined as a 14-day average value

Station 1 Wind Rose – Shows the direction the wind is coming from, the monitoring station being at the center of the rose.





Legend

-  Onsite Fixed Monitoring Locations
-  New-Indy Catawba



Period H₂S Monitoring Hydrogen Sulfide Onsite Monitors

Below are graphs for onsite locations during the current reporting period.


Depending on wind direction, the H₂S measured at the onsite fence line locations may not exit the mill property at reported concentrations. Wind directions from offsite locations, blowing onto mill property, will disperse ambient concentrations to lower levels prior to exiting the plant site.

Winds were coming from a variable direction throughout the day at 1 to 16 mph.

See wind rose diagram with aerial map figure for full wind data during this reporting period.



Station3 - 11/21/2024



Submitted Fenceline H₂S and Met 30-minute Data

30-Minute Avgs	Station 1			Station 2			Station 3		
	H2S	Met		H2S	Met		H2S	Met	
	30min Avg H2S Conc.	30min Avg WS	30min Avg WD	30min Avg H2S Conc.	30min Avg WS	30min Avg WD	30min Avg H2S Conc.	30min Avg WS	30min Avg WD
11/21/2024									
Date / Time	ppb	mph	degrees	ppb	mph	degrees	ppb	mph	degrees
11/21/2024 0:30	AX	9.6	324	0.2	4.6	332	0.2	2.6	329
11/21/2024 1:00	2.1	6.6	317	AX	4.9	332	0.2	3.0	338
11/21/2024 1:30	1.3	4.8	279	0.5	2.1	298	AX	0.6	208
11/21/2024 2:00	0.9	4.2	266	0.2	0.3	300	0.4	0.8	213
11/21/2024 2:30	0.7	3.4	234	0.2	0.3	281	0.2	0.9	216
11/21/2024 3:00	0.7	3.6	210	0.4	0.5	256	0.2	0.6	247
11/21/2024 3:30	0.6	3.3	203	0.5	1.2	223	0.2	0.5	295
11/21/2024 4:00	0.6	3.5	206	0.2	0.7	209	0.7	0.2	176
11/21/2024 4:30	0.6	3.3	191	0.2	0.7	183	1.5	0.6	185
11/21/2024 5:00	0.6	3.0	194	0.2	1.0	191	0.9	0.4	254
11/21/2024 5:30	0.5	3.2	200	0.2	1.4	207	1.0	0.4	202
11/21/2024 6:00	0.6	2.9	198	1.2	1.1	196	1.2	0.5	313
11/21/2024 6:30	0.6	2.6	199	0.2	1.1	196	1.3	0.6	186
11/21/2024 7:00	0.5	2.0	225	0.2	1.3	215	1.0	0.5	249
11/21/2024 7:30	0.6	0.7	153	3.4	0.9	219	1.4	0.3	305
11/21/2024 8:00	0.9	1.8	156	0.2	2.6	188	2.1	0.7	168
11/21/2024 8:30	0.6	3.0	200	0.4	3.1	203	2.8	1.0	179
11/21/2024 9:00	0.7	3.9	202	0.9	3.9	215	2.7	2.6	201
11/21/2024 9:30	0.7	5.0	207	0.7	4.6	214	0.7	2.7	224
11/21/2024 10:00	0.7	5.9	213	0.4	4.2	217	0.6	2.8	220
11/21/2024 10:30	0.7	7.5	232	1.3	6.2	237	0.2	3.4	244
11/21/2024 11:00	0.7	7.8	217	0.8	7.0	234	0.4	3.9	234
11/21/2024 11:30	0.7	7.5	227	0.6	7.2	234	1.3	3.9	224
11/21/2024 12:00	0.7	10.0	219	0.7	8.0	233	1.1	4.5	219
11/21/2024 12:30	0.7	9.3	228	0.9	8.9	242	0.8	3.8	239
11/21/2024 13:00	0.6	11.2	242	0.7	8.7	237	0.2	4.3	242
11/21/2024 13:30	0.6	9.9	242	0.6	9.2	244	0.5	4.3	248
11/21/2024 14:00	0.6	9.1	255	1.5	9.1	255	0.2	3.5	250
11/21/2024 14:30	0.6	10.1	241	1.8	9.1	255	0.2	2.7	276
11/21/2024 15:00	0.6	10.1	244	1.5	9.6	244	0.2	3.2	254
11/21/2024 15:30	0.6	11.0	251	1.4	10.5	248	0.2	3.2	257
11/21/2024 16:00	0.2	6.6	253	2.1	8.3	253	0.2	2.9	262
11/21/2024 16:30	0.2	5.9	239	2.2	6.4	253	0.2	1.6	268
11/21/2024 17:00	0.2	5.6	233	1.8	5.4	246	0.2	1.6	287
11/21/2024 17:30	1.4	5.2	308	0.8	2.4	333	0.2	0.9	306
11/21/2024 18:00	0.8	16.4	324	0.2	8.0	310	0.2	2.3	309
11/21/2024 18:30	0.7	13.1	319	0.2	7.1	306	0.2	2.3	296
11/21/2024 19:00	0.9	9.6	316	0.2	5.1	307	0.2	1.0	234
11/21/2024 19:30	0.8	9.8	326	0.2	3.0	330	0.2	1.8	246
11/21/2024 20:00	0.9	9.9	320	0.2	3.9	311	0.2	1.3	256
11/21/2024 20:30	0.9	5.7	309	0.2	2.2	299	0.2	0.7	213
11/21/2024 21:00	0.7	3.3	291	0.2	1.8	282	0.2	0.6	213
11/21/2024 21:30	0.6	3.4	286	0.2	1.8	290	0.2	0.5	205
11/21/2024 22:00	0.6	2.7	256	0.2	1.2	282	0.2	0.6	196
11/21/2024 22:30	0.4	2.3	210	0.2	0.4	331	0.2	0.8	205
11/21/2024 23:00	0.5	1.5	189	0.2	1.2	280	0.2	0.8	216
11/21/2024 23:30	0.6	2.5	230	1.2	2.0	243	0.2	0.8	231
11/22/2024 0:00	0.6	2.6	192	3.4	1.3	236	0.2	0.8	232

AQS Null Data Codes	
Qualifier Code	Item Description
AB	TECHNICIAN UNAVAILABLE
AC	CONSTRUCTION/REPAIRS IN AREA
AD	SHELTER STORM DAMAGE
AE	SHELTER TEMPERATURE OUTSIDE LIMITS
AI	INSUFFICIENT DATA (CAN'T CALCULATE)
AM	MISCELLANEOUS VOID
AN	MACHINE MALFUNCTION
AO	BAD WEATHER
AP	VANDALISM
AS	POOR QUALITY ASSURANCE RESULTS
AT	CALIBRATION
AU	MONITORING WAIVED
AV	POWER FAILURE (POWR)
AW	WILDLIFE DAMAGE
AX	PRECISION CHECK (PREC)
AY	Q C CONTROL POINTS (ZERO/SPAN)
BA	MAINTENANCE/ROUTINE REPAIRS
BB	UNABLE TO REACH SITE
BC	MULTI-POINT CALIBRATION
BD	AUTO CALIBRATION
BE	BUILDING/SITE REPAIR
BF	PRECISION/ZERO/SPAN
BJ	OPERATOR ERROR
BK	SITE COMPUTER/DATA LOGGER DOWN
BL	Q C AUDIT (AUDT)
EC	EXCEED CRITICAL CRITERIA