

Air Monitoring Summary Tables

The table below summarizes monitoring data collected using the H₂S analyzers deployed at the onsite stations.
All times in Eastern Standard Time (EST).

From: 10/02/26 12:00 am

To: 10/02/26

11:59 pm

Onsite Fenceline Monitors


Instrument	Analyte	30-min AEGL Reached?	Concentration Range Detected ^a	24-hr Average ^a	7-day Average	30-min AEGL
Station 1						
TAPI Analyzer	H ₂ S	No	1 – 3 ppb	1.50 ppb	7.40 ppb	600 ppb
Station 2						
TAPI Analyzer	H ₂ S	No	0 – 11 ppb	1.51 ppb	0.89 ppb	600 ppb
Station 3						
TAPI Analyzer	H ₂ S	No	0 – 18 ppb	3.76 ppb	1.14 ppb	600 ppb

^a Based on 30-minute averages.

Notes:

ATSDR MRL	Agency for Toxic Substances and Disease Registry Minimal Risk Level (MRL)
AEGL	EPA Acute Exposure Guidelines Levels
H ₂ S	Hydrogen Sulfide
TAPI	Teledyne API H ₂ S Analyzer
hr	Hour
min	Minute
ppb	Parts per billion
MRL Limit	Limit defined as a 14-day average value

Station 1 Wind Rose – Shows the direction the wind is coming from, the monitoring station being at the center of the rose.





Legend

-  Onsite Fixed Monitoring Locations
-  New-Indy Catawba



Period H₂S Monitoring Hydrogen Sulfide Onsite Monitors

Below are graphs for onsite locations during the current reporting period.


Depending on wind direction, the H₂S measured at the onsite fence line locations may not exit the mill property at reported concentrations. Wind directions from offsite locations, blowing onto mill property, will disperse ambient concentrations to lower levels prior to exiting the plant site.

Winds were predominantly coming from the southwest, south-southwest, and south direction throughout the day at 1 to 6 mph.

See wind rose diagram with an aerial map figure for full wind data during this reporting period.



Station3 - 10/2/2026



Submitted Fenceline H₂S and Met 30-minute Data

30-Minute Averages	Station 1			Station 2			Station 3		
	Station 1	Met		H2S	Met		H2S	Met	
	30min Avg H2S Conc.	30min Avg WS	30min Avg WD	30min Avg H2S Conc.	30min Avg WS	30min Avg WD	30min Avg H2S Conc.	30min Avg WS	30min Avg WD
Date / Time	ppb	mph	degrees	ppb	mph	degrees	ppb	mph	degrees
10/2/2026 0:30	AX	2.1	206	0.5	0.2	353	1.1	0.4	230
10/2/2026 1:00	2.9	1.3	177	AX	0.2	31	1.1	0.3	39
10/2/2026 1:30	2.2	2.0	193	0.8	0.2	345	AX	0.3	32
10/2/2026 2:00	1.6	1.0	196	1.0	0.2	1	1.1	0.4	23
10/2/2026 2:30	1.5	1.4	192	0.2	0.2	16	0.7	0.3	16
10/2/2026 3:00	1.4	1.3	184	0.2	0.2	30	0.2	0.2	140
10/2/2026 3:30	1.4	1.7	196	0.2	0.2	342	0.2	0.2	155
10/2/2026 4:00	1.4	1.0	190	0.2	0.2	33	0.5	0.2	87
10/2/2026 4:30	1.3	1.9	200	0.2	0.2	346	0.2	0.2	5
10/2/2026 5:00	1.4	1.4	201	0.2	0.2	331	0.2	0.2	134
10/2/2026 5:30	1.4	1.3	184	0.2	0.2	124	0.2	0.2	76
10/2/2026 6:00	1.5	1.4	180	0.2	0.2	139	0.2	0.2	67
10/2/2026 6:30	1.6	1.1	177	0.2	0.2	19	0.2	0.3	60
10/2/2026 7:00	1.5	1.4	178	0.2	0.2	12	0.2	0.3	72
10/2/2026 7:30	1.6	1.8	192	0.2	0.2	339	0.4	0.3	76
10/2/2026 8:00	1.7	1.9	188	0.2	0.2	320	1.2	0.5	169
10/2/2026 8:30	1.6	2.8	185	0.2	0.2	202	4.8	1.2	173
10/2/2026 9:00	1.5	3.0	189	0.2	0.2	199	5.7	2.0	175
10/2/2026 9:30	1.5	2.7	197	0.9	0.2	205	16.1	1.8	149
10/2/2026 10:00	1.4	4.2	243	5.5	3.3	225	3.3	1.9	221
10/2/2026 10:30	1.3	5.6	239	7.8	4.1	237	2.5	2.5	212
10/2/2026 11:00	1.3	4.5	227	4.7	4.0	230	1.0	2.5	214
10/2/2026 11:30	1.3	5.4	224	1.4	3.2	217	1.7	2.3	210
10/2/2026 12:00	1.3	5.2	215	1.9	3.5	226	1.7	3.0	237
10/2/2026 12:30	1.3	5.5	215	1.5	3.3	246	1.1	2.6	214
10/2/2026 13:00	1.3	5.1	250	1.9	3.1	232	0.5	1.8	171
10/2/2026 13:30	1.3	5.2	314	2.8	4.0	255	2.1	3.4	211
10/2/2026 14:00	1.3	5.1	235	1.3	3.9	266	0.9	3.1	213
10/2/2026 14:30	1.3	5.2	263	2.4	3.6	245	0.6	2.4	200
10/2/2026 15:00	1.3	4.2	215	2.2	3.7	227	0.6	2.4	15
10/2/2026 15:30	1.3	2.3	238	2.0	2.7	224	0.4	2.2	238
10/2/2026 16:00	1.3	4.6	233	1.9	3.6	215	0.9	2.1	207
10/2/2026 16:30	1.5	3.2	218	2.1	2.2	218	13.2	1.9	220
10/2/2026 17:00	1.7	3.2	214	2.2	1.7	221	8.7	2.0	208
10/2/2026 17:30	1.4	3.6	221	1.9	2.1	224	18.2	1.4	226
10/2/2026 18:00	2.0	1.7	201	11.2	0.8	192	10.4	0.3	263
10/2/2026 18:30	1.7	2.5	179	2.6	0.8	86	13.6	0.3	191
10/2/2026 19:00	1.6	1.8	204	0.9	0.5	138	11.4	0.3	167
10/2/2026 19:30	1.3	2.9	209	0.9	0.2	257	7.5	0.2	149
10/2/2026 20:00	1.3	2.4	211	0.2	0.2	103	4.2	0.4	128
10/2/2026 20:30	1.4	2.6	211	0.2	0.3	286	3.7	0.4	292
10/2/2026 21:00	1.4	2.7	203	0.2	0.3	102	2.9	0.2	100
10/2/2026 21:30	1.4	2.1	213	0.9	0.4	85	2.1	0.5	7
10/2/2026 22:00	1.8	3.7	184	1.7	0.8	15	2.6	0.6	67
10/2/2026 22:30	1.4	2.3	200	1.3	0.8	64	2.7	0.7	7
10/2/2026 23:00	1.6	2.9	209	1.0	0.4	181	3.4	0.6	353
10/2/2026 23:30	1.6	3.4	194	0.4	0.6	176	7.6	0.6	148
10/3/2026 0:00	1.6	2.8	203	0.2	0.5	166	13.0	0.7	189

AQS Null Data Codes	
Qualifier Code	Item Description
AB	TECHNICIAN UNAVAILABLE
AC	CONSTRUCTION/REPAIRS IN AREA
AD	SHELTER STORM DAMAGE
AE	SHELTER TEMPERATURE OUTSIDE LIMITS
AI	INSUFFICIENT DATA (CAN'T CALCULATE)
AM	MISCELLANEOUS VOID
AN	MACHINE MALFUNCTION
AO	BAD WEATHER
AP	VANDALISM
AS	POOR QUALITY ASSURANCE RESULTS
AT	CALIBRATION
AU	MONITORING WAIVED
AV	POWER FAILURE (POWR)
AW	WILDLIFE DAMAGE
AX	PRECISION CHECK (PREC)
AY	QC CONTROL POINTS (ZERO/SPAN)
BA	MAINTENANCE/ROUTINE REPAIRS
BB	UNABLE TO REACH SITE
BC	MULTI-POINT CALIBRATION
BD	AUTO CALIBRATION
BE	BUILDING/SITE REPAIR
BF	PRECISION/ZERO/SPAN
BJ	OPERATOR ERROR
BK	SITE COMPUTER/DATA LOGGER DOWN
BL	QC AUDIT (AUDT)
EC	EXCEED CRITICAL CRITERIA